

Temperature Log for Freezer - Fahrenheit Days 1-15

November 2020 | tempcheck@phila.gov

Month/Year _____ Facility Name _____

If you have a vaccine storage issue, contact us immediately!

VFC or VFAAR Pin _____ Room _____

Storage and Handling: TempCheck@phila.gov, 215-685-6777

VFC/VFAAR Coordinator: 215-685-6424

Day of Month	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
Staff Initials	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Exact Time	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm	am • pm
	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Min/Max Temp (since previous reading)
Danger! Temperatures above 5°F are too warm! Write any out-of-range temps and room temp on the lines below and call the Immunization Program immediately!															
in-range temperatures	5°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	4°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	3°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	2°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	1°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	0°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-1°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-2°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-3°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-4°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-58°F to -5°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:
action	Write any out-of-range temps (above 5°F or below -58°F) here.	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	Room Temperature	:	:	:	:	:	:	:	:	:	:	:	:	:	:

Monitor Temperatures Closely!

1. Monitor storage temperatures with a certified, calibrated digital data logger.
2. Write your initials below in "Staff Initials," and note the time in "Exact Time."
3. Record temps twice each workday.
4. Record the min/max temps once each workday – preferably in the morning.
5. Put an "X" in the row that corresponds to the refrigerator's temperature.
6. If any out-of-range temp, see instructions to the right.
7. After each month has ended, save each month's log for 3 years.

Take Action if Temperature is Out of Range!

Too warm (above -15°C) or too cold (below -50°C).

1. Consult your **Emergency Management Plan**.
2. Label exposed vaccine "do not use," and store it under proper conditions as quickly as possible. Do not discard vaccines unless directed by the Philadelphia Immunization Program.
3. Record the out-of-range temps and the room temp in the "Action" area on the bottom of the log.
4. Notify your vaccine coordinator and contact TempCheck (TempCheck@phila.gov, 215-685-6777).
5. Document all action taken.

Temperature Log for Freezer - Fahrenheit Days 16-31

November 2020 | tempcheck@phila.gov

Month/Year _____ Facility Name _____

If you have a vaccine storage issue, contact us immediately!

VFC or VFAAR Pin _____ Room _____

Storage and Handling: TempCheck@phila.gov, 215-685-6777

VFC/VFAAR Coordinator: 215-685-6424

Day of Month	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Staff Initials	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Exact Time	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm	am : pm
	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
Min/Max Temp (since previous reading)
Danger! Temperatures above 5°F are too warm! Write any out-of-range temps and room temp on the lines below and call the Immunization Program immediately!																
in-range temperatures	5°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	4°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	3°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	2°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	1°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	0°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-1°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-2°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-3°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	-4°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
-58°F to -5°F	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
action	Write any out-of-range temps (above 5°F or below -58°F) here.	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:
	Room Temperature	:	:	:	:	:	:	:	:	:	:	:	:	:	:	:

Monitor Temperatures Closely!

1. Monitor storage temperatures with a certified, calibrated digital data logger.
2. Write your initials below in "Staff Initials," and note the time in "Exact Time."
3. Record temps twice each workday.
4. Record the min/max temps once each workday – preferably in the morning.
5. Put an "X" in the row that corresponds to the refrigerator's temperature.
6. If any out-of-range temp, see instructions to the right.
7. After each month has ended, save each month's log for 3 years.

Take Action if Temperature is Out of Range!

Too warm (above 5°F) or too cold (below -58°F).

1. Consult your **Emergency Management Plan**.
2. Label exposed vaccine "do not use," and store it under proper conditions as quickly as possible. Do not discard vaccines unless directed by the Philadelphia Immunization Program.
3. Record the out-of-range temps and the room temp in the "Action" area on the bottom of the log.
4. Notify your vaccine coordinator and contact TempCheck (TempCheck@phila.gov, 215-685-6777).
5. Document all action taken.